Interferential Therapy Machines in Community-based Services

In laboratory and clinical situations, this treatment modality has resulted in transmission of microorganisms from one individual to another. To reduce the risk of transmission:

- Always clean and disinfect suction cups, sponges and cables between clients according to manufacturer instructions.
- Disposable electrodes are recommended.
- Perform treatment only on clients with intact skin.

Original date: January 2016

Revised date: March 2020